

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXD-0007
Location: KEYPORT POLICE
Calibration File No.: 00513 Calib. Date: 09/17/2013 Calib. No.: 00016
Certification File No.: 00514 Cert. Date: 09/17/2013 Cert. No.: 00012
Linearity File No.: 00515 Lin. Date: 09/17/2013 Lin. No.: 00012
Solution File No.: 00531 Soln. Date: 12/20/2013 Soln. No.: 00117
Sequential File No.: 00531 File Date: 12/20/2013

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDZK-0100
Control Solution %: 0.100% Expires: 06/17/2015
Solution Control Lot: 13F114 Bottle No.: 0628

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:25S		
Control 1 EC	0.100%	08:26S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	08:26S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:27S		
Control 2 EC	0.099%	08:27S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:27S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:28S		
Control 3 EC	0.100%	08:29S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	08:29S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:30S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: AUMACK

First Name: ROBERT

MI: L

Signature: _____

Badge No.: 19

Date: 12/20/2013

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXE-0082
Location: UNION BEACH POLICE
Calibration File No.: 00312 Calib. Date: 08/20/2013 Calib. No.: 00015
Certification File No.: 00313 Cert. Date: 08/20/2013 Cert. No.: 00011
Linearity File No.: 00314 Lin. Date: 08/20/2013 Lin. No.: 00011
Solution File No.: 00315 Soln. Date: 08/20/2013 Soln. No.: 00084
Sequential File No.: 00315 File Date: 08/20/2013

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDZL-0158
Control Solution %: 0.100% Expires: 11/05/2014
Solution Control Lot: 12K108 Bottle No.: 0105

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:14D		
Control 1 EC	0.100%	16:14D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	16:14D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:15D		
Control 2 EC	0.099%	16:15D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	16:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:16D		
Control 3 EC	0.098%	16:17D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	16:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:17D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDXA P2-231 (10)

Changed By:

Last Name: DENNIS

First Name: MARC

MI: W.

Signature: SGT. [Signature] #5925

Badge No.: 5925

Date: 08/20/2013

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXD-0007
Location:	KEYPORT POLICE		
Calibration File No.:	00482	Calib. Date:	03/19/2013
Certification File No.:	00483	Cert. Date:	03/19/2013
Linearity File No.:	00484	Lin. Date:	03/19/2013
Solution File No.:	00509	Soln. Date:	08/30/2013
Sequential File No.:	00509	File Date:	08/30/2013
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDZK-0100
Solution Control Lot:	12K108	Expires:	11/05/2014
		Bottle No.:	0622

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:54D		
Control 1 EC	0.100%	08:55D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	08:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:55D		
Control 2 EC	0.100%	08:56D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	08:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:57D		
Control 3 EC	0.099%	08:57D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	08:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:58D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: AUMACK

First Name: ROBERT

MI: L

Signature: 

Badge No.: 19

Date: 08/30/2013

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXD-0007
Location:	KEYPORT POLICE		
Calibration File No.:	00455	Calib. Date:	09/25/2012
Certification File No.:	00456	Cert. Date:	09/25/2012
Linearity File No.:	00457	Lin. Date:	09/25/2012
Solution File No.:	00481	Soln. Date:	03/12/2013
Sequential File No.:	00481	File Date:	03/12/2013
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDZK-0101
Solution Control Lot:	111095	Expires:	09/15/2013
		Bottle No.:	0148

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:55S		
Control 1 EC	0.099%	07:55S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	07:55S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:56S		
Control 2 EC	0.098%	07:57S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	07:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:58S		
Control 3 EC	0.098%	07:58S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	07:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:59S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ELY

First Name: RICHARD

MI: B

Signature: _____

Handwritten signature: R. Ely #12

Badge No.: 12

Date: 03/12/2013

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXD-0007
Location:	KEYPORT POLICE		
Calibration File No.:	00455	Calib. Date:	09/25/2012
Certification File No.:	00456	Cert. Date:	09/25/2012
Linearity File No.:	00457	Lin. Date:	09/25/2012
Solution File No.:	00478	Soln. Date:	03/01/2013
Sequential File No.:	00478	File Date:	03/01/2013
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDZK-0101
Solution Control Lot:	11I095	Expires:	09/15/2013
		Bottle No.:	0147

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	03:33S		
Control 1 EC	0.097%	03:34S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	03:34S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:35S		
Control 2 EC	0.097%	03:35S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	03:35S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:36S		
Control 3 EC	0.097%	03:37S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	03:37S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:37S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: FERM

First Name: MICHAEL

MI: A

Signature: _____

SGT. Michael F. Fer #16

Badge No.: 2216

Date: 03/01/2013

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXD-0007
Location: KEYPORT POLICE
Calibration File No.: 00455 Calib. Date: 09/25/2012 Calib. No.: 00014
Certification File No.: 00456 Cert. Date: 09/25/2012 Cert. No.: 00010
Linearity File No.: 00457 Lin. Date: 09/25/2012 Lin. No.: 00010
Solution File No.: 00475 Soln. Date: 01/28/2013 Soln. No.: 00103
Sequential File No.: 00475 File Date: 01/28/2013

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDZK-0101
Control Solution %: 0.100% Expires: 09/15/2013
Solution Control Lot: 11I095 Bottle No.: 0146

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:21S		
Control 1 EC	0.100%	12:22S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	12:22S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:22S		
Control 2 EC	0.099%	12:23S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	12:23S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:24S		
Control 3 EC	0.099%	12:24S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	12:24S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:25S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ELY

First Name: RICHARD

MI: B

Signature: _____

Badge No.: 12

Date: 01/28/2013