

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXD-0007
Location:	KEYPORT POLICE		
Calibration File No.:	00851	Calib. Date:	03/27/2019
Certification File No.:	00852	Cert. No.:	00029
Linearity File No.:	00853	Cert. No.:	00024
Solution File No.:	00887	Lin. Date:	03/27/2019
Sequential File No.:	00887	Soln. No.:	00024
		Soln. No.:	00209
		File Date:	08/07/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDZK-0101
Solution Control Lot:	17290	Expires:	09/14/2019
		Bottle No.:	1397

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	21:11D		
Control 1 EC	0.098%	21:11D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	21:11D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:12D		
Control 2 EC	0.099%	21:13D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	21:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:13D		
Control 3 EC	0.098%	21:14D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	21:14D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:14D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HASSMILLER

First Name: ROBERT

MI: J

Signature: 

Badge No.: 2232

Date: 08/07/2019