

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXD-0007
Location: KEYPORT POLICE
Calibration File No.: 00851 **Calib. Date:** 03/27/2019 **Calib. No.:** 00029
Certification File No.: 00852 **Cert. Date:** 03/27/2019 **Cert. No.:** 00024
Linearity File No.: 00853 **Lin. Date:** 03/27/2019 **Lin. No.:** 00024
Solution File No.: 00879 **Soln. Date:** 06/24/2019 **Soln. No.:** 00206
Sequential File No.: 00879 **File Date:** 06/24/2019

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDZK-0101
Control Solution %: 0.100% **Expires:** 09/14/2019
Solution Control Lot: 17290 **Bottle No.:** 1375

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	01:21D		
Control 1 EC	0.097%	01:21D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	01:21D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:22D		
Control 2 EC	0.098%	01:22D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	01:22D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:23D		
Control 3 EC	0.098%	01:24D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	01:24D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:24D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HASSMILLER

First Name: ROBERT

MI: J

Signature: 

Badge No.: 2232

Date: 06/24/2019