

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXD-0007  
**Location:** KEYPORT POLICE  
**Calibration File No.:** 00647 **Calib. Date:** 01/27/2016 **Calib. No.:** 00021  
**Certification File No.:** 00648 **Cert. Date:** 01/27/2016 **Cert. No.:** 00017  
**Linearity File No.:** 00649 **Lin. Date:** 01/27/2016 **Lin. No.:** 00017  
**Solution File No.:** 00662 **Soln. Date:** 04/09/2016 **Soln. No.:** 00151  
**Sequential File No.:** 00662 **File Date:** 04/09/2016

**Calibrating Unit:** WET **Model No.:** CU-34 **Serial No.:** DDZK-0101  
**Control Solution %:** 0.100% **Expires:** 12/31/2016  
**Solution Control Lot:** 14L135 **Bottle No.:** 1226

| Function          | Result<br>%BAC | Time<br>HH:MM | Temperature<br>Simulator (°C) | Comment(s)<br>or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000%         | 21:51D        |                               |                           |
| Control 1 EC      | 0.099%         | 21:52D        | 34.0°C                        | *** TEST PASSED ***       |
| Control 1 IR      | 0.099%         | 21:52D        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 21:53D        |                               |                           |
| Control 2 EC      | 0.099%         | 21:53D        | 34.0°C                        | *** TEST PASSED ***       |
| Control 2 IR      | 0.099%         | 21:53D        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 21:54D        |                               |                           |
| Control 3 EC      | 0.098%         | 21:55D        | 34.0°C                        | *** TEST PASSED ***       |
| Control 3 IR      | 0.099%         | 21:55D        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 21:56D        |                               |                           |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: WHEELER

First Name: STEPHEN

MI: C

Signature: ITL. JECR #27

Badge No.: 27

Date: 04/09/2016