

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXD-0007
Location: KEYPORT POLICE
Calibration File No.: 00620 **Calib. Date:** 08/14/2015 **Calib. No.:** 00020
Certification File No.: 00621 **Cert. Date:** 08/14/2015 **Cert. No.:** 00016
Linearity File No.: 00622 **Lin. Date:** 08/14/2015 **Lin. No.:** 00016
Solution File No.: 00643 **Soln. Date:** 12/27/2015 **Soln. No.:** 00147
Sequential File No.: 00643 **File Date:** 12/27/2015

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDZK-0100
Control Solution %: 0.100% **Expires:** 12/31/2016
Solution Control Lot: 14L135 **Bottle No.:** 1236

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:07S		
Control 1 EC	0.097%	21:08S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	21:08S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:08S		
Control 2 EC	0.098%	21:09S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	21:09S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:10S		
Control 3 EC	0.098%	21:10S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	21:10S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:11S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WHEELER

First Name: STEPHEN

MI: C

Signature: _____

FTL. JEWELL #27

Badge No.: 27

Date: 12/27/2015