Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARXD-0007
Location:	KEYPORT POLICE				
Calibration File No.:	00675	Calib. Date:	07/05/2016	Calib. No.:	00022
Certification File No.:	00676	Cert. Date:	07/05/2016	Cert. No.:	00018
Linearity File No.:	00677	Lin. Date:	07/05/2016	Lin. No.:	00018
Solution File No.:	00700	Soln. Date:	10/19/2016	Soln. No.:	00160
Sequential File No.:	00700	File Date:	10/19/2016		
Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDZK-0101
Control Solution %:	0.100%			Expires:	12/31/2016
Solution Control Lot:	14L135			Bottle No.:	1227


Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	05:23D		
Control 1 EC	0.099%	05:24D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	05:24D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:25D		
Control 2 EC	0.097%	05:25D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	05:25D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:26D		
Control 3 EC	0.096%	05:27D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	05:27D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:28D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SALVATORE First Name: SGT. MATTHEW MI: -

Signature:  Badge No.: 2221

Date: 10/19/2016