

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXD-0007
Location:	KEYPORT POLICE		
Calibration File No.:	00705	Calib. Date:	12/06/2016
Certification File No.:	00706	Cert. Date:	12/06/2016
Linearity File No.:	00707	Lin. Date:	12/06/2016
Solution File No.:	00730	Soln. Date:	04/29/2017
Sequential File No.:	00730	File Date:	04/29/2017
		Calib. No.:	00023
		Cert. No.:	00019
		Lin. No.:	00019
		Soln. No.:	00166
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDZK-0100
Solution Control Lot:	16330	Expires:	10/28/2018
		Bottle No.:	1479

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:52D		
Control 1 EC	0.102%	02:53D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	02:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:54D		
Control 2 EC	0.101%	02:54D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	02:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:55D		
Control 3 EC	0.101%	02:56D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	02:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:57D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: RUTH

First Name: JOSEPH

MI: H

Signature:



Badge No.: 2224

Date: 04/29/2017