

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXD-0007
Location: KEYPORT POLICE
Calibration File No.: 00647 **Calib. Date:** 01/27/2016 **Calib. No.:** 00021
Certification File No.: 00648 **Cert. Date:** 01/27/2016 **Cert. No.:** 00017
Linearity File No.: 00649 **Lin. Date:** 01/27/2016 **Lin. No.:** 00017
Solution File No.: 00671 **Soln. Date:** 06/12/2016 **Soln. No.:** 00153
Sequential File No.: 00671 **File Date:** 06/12/2016

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDZK-0101
Control Solution %: 0.100% **Expires:** 12/31/2016
Solution Control Lot: 14L135 **Bottle No.:** 1238

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	04:11D		
Control 1 EC	0.100%	04:12D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	04:12D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:13D		
Control 2 EC	0.100%	04:13D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	04:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:14D		
Control 3 EC	0.100%	04:15D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	04:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:15D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SALVATORE

First Name: SGT. MATTHEW SALVATORE

MI: -

Signature: _____

Badge No.: 2221

Date: 06/12/2016