

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXD-0007
Location:	KEYPORT POLICE		
Calibration File No.:	00732	Calib. Date:	05/16/2017
Certification File No.:	00733	Cert. Date:	05/16/2017
Linearity File No.:	00734	Lin. Date:	05/16/2017
Solution File No.:	00761	Soln. Date:	08/20/2017
Sequential File No.:	00761	File Date:	08/20/2017
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDZK-0101
Solution Control Lot:	16330	Expires:	10/28/2018
		Bottle No.:	1417

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	23:04D		
Control 1 EC	0.098%	23:05D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	23:05D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:06D		
Control 2 EC	0.098%	23:06D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	23:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:07D		
Control 3 EC	0.097%	23:08D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	23:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:09D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: RUTH

First Name: JOSEPH

MI: H

Signature: _____

Sgt J Ruth #24

Badge No.: 2224

Date: 08/20/2017