

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXD-0007
Location: KEYPORT POLICE
Calibration File No.: 00732 **Calib. Date:** 05/16/2017 **Calib. No.:** 00024
Certification File No.: 00733 **Cert. Date:** 05/16/2017 **Cert. No.:** 00020
Linearity File No.: 00734 **Lin. Date:** 05/16/2017 **Lin. No.:** 00020
Solution File No.: 00768 **Soln. Date:** 09/06/2017 **Soln. No.:** 00179
Sequential File No.: 00768 **File Date:** 09/06/2017

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDZK-0101
Control Solution %: 0.100% **Expires:** 10/28/2018
Solution Control Lot: 16330 **Bottle No.:** 1414

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	05:58D		
Control 1 EC	0.096%	05:59D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.096%	05:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:59D		
Control 2 EC	0.095%	06:00D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.096%	06:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:01D		
Control 3 EC	0.095%	06:02D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	06:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:02D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SALVATORE

First Name: SGT. MATTHEW

MI: -

Signature: 

Badge No.: 2221

Date: 09/06/2017