

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXD-0007
Location:	KEYPORT POLICE		
Calibration File No.:	00732	Calib. Date:	05/16/2017
Certification File No.:	00733	Cert. Date:	05/16/2017
Linearity File No.:	00734	Lin. Date:	05/16/2017
Solution File No.:	00768	Soln. Date:	09/06/2017
Sequential File No.:	00768	File Date:	09/06/2017
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDZK-0101
Solution Control Lot:	16330	Expires:	10/28/2018
		Bottle No.:	1414

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	05:58D		
Control 1 EC	0.096%	05:59D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.096%	05:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:59D		
Control 2 EC	0.095%	06:00D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.096%	06:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:01D		
Control 3 EC	0.095%	06:02D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	06:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:02D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SALVATORE

First Name: SGT. MATTHEW

MI: -

Signature: _____

Badge No.: 2221

Date: 09/06/2017