

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXD-0007
Location:	KEYPORT POLICE		
Calibration File No.:	00620	Calib. Date:	08/14/2015
Certification File No.:	00621	Cert. Date:	08/14/2015
Linearity File No.:	00622	Lin. Date:	08/14/2015
Solution File No.:	00634	Soln. Date:	09/06/2015
Sequential File No.:	00634	File Date:	09/06/2015
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDZK-0100
Solution Control Lot:	14L135	Expires:	12/31/2016
		Bottle No.:	1250

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	22:41D		
Control 1 EC	0.097%	22:41D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	22:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:42D		
Control 2 EC	0.096%	22:43D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.096%	22:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:43D		
Control 3 EC	0.096%	22:44D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.096%	22:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:45D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SALVATORE

First Name: SGT. MATTHEW

MI: -

Signature: _____

Badge No.: 2221

Date: 09/06/2015

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXD-0007
Location: KEYPORT POLICE
Calibration File No.: 00620 Calib. Date: 08/14/2015 Calib. No.: 00020
Certification File No.: 00621 Cert. Date: 08/14/2015 Cert. No.: 00016
Linearity File No.: 00622 Lin. Date: 08/14/2015 Lin. No.: 00016
Solution File No.: 00636 Soln. Date: 09/07/2015 Soln. No.: 00144
Sequential File No.: 00636 File Date: 09/07/2015

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDZK-0100
Control Solution %: 0.100% Expires: 12/31/2016
Solution Control Lot: 14L135 Bottle No.: 1242

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	04:14D		
Control 1 EC	0.098%	04:15D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	04:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:15D		
Control 2 EC	0.098%	04:16D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	04:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:17D		
Control 3 EC	0.098%	04:17D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	04:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:18D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WHEELER

First Name: STEPHEN

MI: C

Signature: _____

PTZ. JECW #27

Badge No.: 27

Date: 09/07/2015

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXD-0007
Location:	KEYPORT POLICE		
Calibration File No.:	00620	Calib. Date:	08/14/2015
Certification File No.:	00621	Cert. Date:	08/14/2015
Linearity File No.:	00622	Lin. Date:	08/14/2015
Solution File No.:	00626	Soln. Date:	08/30/2015
Sequential File No.:	00626	File Date:	08/30/2015
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDZK-0100
Solution Control Lot:	14L135	Expires:	12/31/2016
		Bottle No.:	1249

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	06:07D		
Control 1 EC	0.098%	06:08D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	06:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:09D		
Control 2 EC	0.098%	06:09D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	06:09D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:10D		
Control 3 EC	0.098%	06:11D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	06:11D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:11D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: RUTH

First Name: JOSEPH

MI: H

Signature:



Badge No.: 2224

Date: 08/30/2015

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXD-0007
Location:	KEYPORT POLICE		
Calibration File No.:	00588	Calib. Date:	02/27/2015
Certification File No.:	00589	Cert. Date:	02/27/2015
Linearity File No.:	00590	Lin. Date:	02/27/2015
Solution File No.:	00619	Soln. Date:	07/28/2015
Sequential File No.:	00619	File Date:	07/28/2015
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDZK-0100
Solution Control Lot:	14L135	Expires:	12/31/2016
		Bottle No.:	1247

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:52D		
Control 1 EC	0.101%	12:53D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	12:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:54D		
Control 2 EC	0.100%	12:54D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	12:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:55D		
Control 3 EC	0.100%	12:56D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	12:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:56D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DENNIS

First Name: MARC

MI: W.

Signature: _____

SGT. B. Dennis #5925

Badge No.: 5925

Date: 07/28/2015

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXD-0007
Location:	KEYPORT POLICE		
Calibration File No.:	00588	Calib. Date:	02/27/2015
Certification File No.:	00589	Cert. Date:	02/27/2015
Linearity File No.:	00590	Lin. Date:	02/27/2015
Solution File No.:	00613	Soln. Date:	07/01/2015
Sequential File No.:	00613	File Date:	07/01/2015
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDZK-0100
Solution Control Lot:	14L135	Expires:	12/31/2016
		Bottle No.:	1246

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:36D		
Control 1 EC	0.102%	09:37D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	09:37D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:38D		
Control 2 EC	0.102%	09:38D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:39D		
Control 3 EC	0.101%	09:40D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	09:40D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:40D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: RUTH

First Name: JOSEPH

MI: H

Signature: _____

Joseph Ruth #24

Badge No.: 2224

Date: 07/01/2015

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: KEYPORT POLICE
Calibration File No.: 00588
Certification File No.: 00589
Linearity File No.: 00590
Solution File No.: 00610
Sequential File No.: 00610

Serial No.: ARXD-0007

Calib. Date: 02/27/2015
Cert. Date: 02/27/2015
Lin. Date: 02/27/2015
Soln. Date: 05/11/2015
File Date: 05/11/2015

Calib. No.: 00019
Cert. No.: 00015
Lin. No.: 00015
Soln. No.: 00138

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 14L135

Model No.: CU-34

Serial No.: DDZK-0100
Expires: 12/31/2016
Bottle No.: 1248

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:04D		
Control 1 EC	0.104%	19:05D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.103%	19:05D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:06D		
Control 2 EC	0.103%	19:06D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	19:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:07D		
Control 3 EC	0.102%	19:08D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	19:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:08D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number:

DDXA P2-252 (10)

Changed By:

Last Name: DENNIS

First Name: MARC


MI: W.

Signature:

SGT. B. Dennis #5925

Badge No.: 5925

Date: 05/11/2015



Dräger

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.
(F.R. Vol. 59 No. 249 12/19/94 Notices)

Draeger Safety Diagnostics, Inc.

☒ Model: ALCOTEST® CU34

☐ Model: MARK IIA

☐ Other: _____

Serial Number:

DDZK-0100

Certification Date

3-12-15

Technician

BC

Re-Certification Due Date

3-12-16

Dräger

ALCOTEST® 7110 TEMPERATURE PROBE

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest® 7110 Temperature Probe has been tested for accuracy with instrumentation that is traceable to the National Institute of Standards and Technology (NIST).

The manufacturer recommends accuracy verification of the Temperature Probe, within 12 months of the certification date below, or sooner, according to your State Specification.

For accurate temperature readings, the probe value on this certificate, noted below, must be programmed into the Alcotest® 7110.

Serial Number Temp. Probe

DDXA P2-252

Certification date:

3-12-15

Next Certification due:

3-12-16

Probe Value

106

Draeger Safety Diagnostics, Inc.
Technical Service Department

BC

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXD-0007
Location: KEYPORT POLICE
Calibration File No.: 00588 **Calib. Date:** 02/27/2015 **Calib. No.:** 00019
Certification File No.: 00589 **Cert. Date:** 02/27/2015 **Cert. No.:** 00015
Linearity File No.: 00590 **Lin. Date:** 02/27/2015 **Lin. No.:** 00015
Solution File No.: 00610 **Soln. Date:** 05/11/2015 **Soln. No.:** 00138
Sequential File No.: 00610 **File Date:** 05/11/2015

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDZK-0100
Control Solution %: 0.100% **Expires:** 12/31/2016
Solution Control Lot: 14L135 **Bottle No.:** 1248

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	19:04D		
Control 1 EC	0.104%	19:05D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.103%	19:05D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:06D		
Control 2 EC	0.103%	19:06D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	19:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:07D		
Control 3 EC	0.102%	19:08D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	19:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:08D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDXA P2-252 (10)

Changed By:

Last Name: DENNIS

First Name: MARC

MI: W.

Signature: SGT. B. Dennis #5925

Badge No.: 5925

Date: 05/11/2015

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXD-0007
Location:	KEYPORT POLICE		
Calibration File No.:	00588	Calib. Date:	02/27/2015
Certification File No.:	00589	Cert. Date:	02/27/2015
Linearity File No.:	00590	Lin. Date:	02/27/2015
Solution File No.:	00609	Soln. Date:	04/26/2015
Sequential File No.:	00609	File Date:	04/26/2015
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDZK-0101
Solution Control Lot:	13F114	Expires:	06/17/2015
		Bottle No.:	0650

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:42D		
Control 1 EC	0.101%	07:43D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	07:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:44D		
Control 2 EC	0.100%	07:44D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	07:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:45D		
Control 3 EC	0.101%	07:46D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	07:46D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:47D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: RUTH

First Name: JOSEPH

MI: H

Signature: _____

Pat J. Ruth #24

Badge No.: 2224

Date: 04/26/2015

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXD-0007
Location: KEYPORT POLICE
Calibration File No.: 00588 **Calib. Date:** 02/27/2015 **Calib. No.:** 00019
Certification File No.: 00589 **Cert. Date:** 02/27/2015 **Cert. No.:** 00015
Linearity File No.: 00590 **Lin. Date:** 02/27/2015 **Lin. No.:** 00015
Solution File No.: 00601 **Soln. Date:** 03/25/2015 **Soln. No.:** 00136
Sequential File No.: 00601 **File Date:** 03/25/2015

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDZK-0101
Control Solution %: 0.100% **Expires:** 06/17/2015
Solution Control Lot: 13F114 **Bottle No.:** 0649

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:48S		
Control 1 EC	0.102%	15:48S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	15:48S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:49S		
Control 2 EC	0.101%	15:50S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	15:50S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:51S		
Control 3 EC	0.101%	15:51S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	15:51S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:52S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SALVATORE

First Name: PTL MATTHEW

MI: -

Signature:  21

Badge No.: 2221

Date: 03/25/2015

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXD-0007
Location: KEYPORT POLICE
Calibration File No.: 00588 **Calib. Date:** 02/27/2015 **Calib. No.:** 00019
Certification File No.: 00589 **Cert. Date:** 02/27/2015 **Cert. No.:** 00015
Linearity File No.: 00590 **Lin. Date:** 02/27/2015 **Lin. No.:** 00015
Solution File No.: 00591 **Soln. Date:** 02/27/2015 **Soln. No.:** 00135
Sequential File No.: 00591 **File Date:** 02/27/2015

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDZK-0101
Control Solution %: 0.100% **Expires:** 06/17/2015
Solution Control Lot: 13F114 **Bottle No.:** 0647

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:28S		
Control 1 EC	0.101%	16:29S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	16:29S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:30S		
Control 2 EC	0.101%	16:30S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	16:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:31S		
Control 3 EC	0.101%	16:32S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	16:32S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:33S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDXA P2-251 (10)

Changed By:

Last Name: DENNIS

First Name: MARC

MI: W.

Signature: SET. [Signature] #5925

Badge No.: 5925

Date: 02/27/2015

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXD-0007
Location:	KEYPORT POLICE		
Calibration File No.:	00570	Calib. Date:	09/03/2014
Certification File No.:	00571	Cert. Date:	09/03/2014
Linearity File No.:	00572	Lin. Date:	09/03/2014
Solution File No.:	00587	Soln. Date:	02/06/2015
Sequential File No.:	00587	File Date:	02/06/2015
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDZK-0101
Solution Control Lot:	13F114	Expires:	06/17/2015
		Bottle No.:	0646

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:13S		
Control 1 EC	0.101%	09:14S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:14S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:15S		
Control 2 EC	0.099%	09:15S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	09:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:16S		
Control 3 EC	0.099%	09:17S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	09:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:18S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SALVATORE

First Name: PTL. MATTHEW

MI: -

Signature: _____



Badge No.: 2221

Date: 02/06/2015

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXD-0007
Location: KEYPORT POLICE
Calibration File No.: 00570 **Calib. Date:** 09/03/2014 **Calib. No.:** 00018
Certification File No.: 00571 **Cert. Date:** 09/03/2014 **Cert. No.:** 00014
Linearity File No.: 00572 **Lin. Date:** 09/03/2014 **Lin. No.:** 00014
Solution File No.: 00585 **Soln. Date:** 01/03/2015 **Soln. No.:** 00133
Sequential File No.: 00585 **File Date:** 01/03/2015

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDZK-0101
Control Solution %: 0.100% **Expires:** 06/17/2015
Solution Control Lot: 13F114 **Bottle No.:** 0645

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	03:02S		
Control 1 EC	0.100%	03:02S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	03:02S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:03S		
Control 2 EC	0.100%	03:04S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	03:04S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:05S		
Control 3 EC	0.099%	03:05S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	03:05S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:06S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: RUTH

First Name: JOSEPH

MI: H

Signature: 

Badge No.: 2224

Date: 01/03/2015